

Changing the Battery in the Digital Multi-Day Max/Min Thermometer

Field Guide

Task

Change the battery in the digital Max/Min Thermometer.

What You Need

- ☐ A new AA-size battery
- ☐ A small Phillips head screwdriver

In the Field

1. The battery is in the battery compartment in the lower section of the instrument casing.
2. Remove the two little screws located at the upper corners of the compartment cover and lift off the cover.
3. Change the battery, taking care to ensure correct polarity (negative end of battery contacting the spring).
4. Replace the compartment cover and secure with the two screws.
5. Recalibrate the sensors following the *Digital Multi-Day Soil Thermometer Calibration Field Guide*.
6. Reset the instrument using the *Digital Multi-Day Max/Min Thermometer Reset Field Guide*.

